

H-1240DA & H-1240DA.4F

Digital Penetrometer with Automatic Digital Timer



Introduction

The Humboldt Digital Penetrometer with Automatic Digital Timer is of rugged construction, but requires careful handling to safeguard its accuracy. Keep the instrument clean at all times – especially the drop and indicator plungers, which constitute the spindle assembly. Use denatured alcohol as the cleaner for these surfaces.

For complete test procedure(s), refer to the following specifications: ASTM D5, D217, D937, D1168, D1321, D1403, D1831, and AASHTO T49.

Installation

- 1. Level the instrument carefully by means of the leveling screws on the base, sighting the spirit level from above, in order to obtain accurate alignment. After the instrument has been leveled, avoid shifting its location.
- 2. Position the height of the supporting casting by turning the knurled travel nut on the main support post and then tighten the black knob.
- 3. If additional weights are used, these should be taken from the side post and set on the end of the plunger (opposite end of the needle chuck).
- 4. Insert penetrating needle or cone into chuck and tighten the chuck screw.

Instruction

As supplied by Humboldt the unit is programmed to provide a solenoid actuation period (plunger release) of 5 seconds. The panel-mounted, electronic timer uses 120 VAC, 50/60 Hz. In order to use 240 VAC, a step-down transformer or converter suitable for digital circuits is required. The H-1240D.4F is supplied with an H-1042 transformer for this purpose and is plug compatible with the penetrometer cap-plug, but requires the input line cord to the transformer to be wired for local service.

For further timer information refer to the ATC Model 365B Long Range Timer Instructions.

Operation

- 1. Holding the plunger with one hand, squeeze the lever on the front of the support casting with the other to allow raising the plunger & needle to clear the sample. Flip the toggle switch on the front panel to the up position and pre-select the specified time of penetration by turning the knobs on the timer bezel.
- 2. To conduct a test set the needle, cone or micro-cone using the shadow of its tip to just meet the sample surface, and then lower the indicator spindle to rest on the top of the plunger. Press the indicator ON/clr button and .0 will appear on the display. When ready to begin, press the red pushbutton under the timer on the front panel to release the plunger for the pre-selected amount time. The timer-controlled electromagnet will immediately release the plunger and after the pre-selected time interval

- will re-grip it to stop the test.
- 3. Measure the travel of the plunger assembly by lowering the digital indicator spindle to contact the plunger and observing the indicator reading.
- 4. Upon conclusion of the test squeeze the lever to release the plunger and raise it and the indicator spindle. To reset the instrument for the next test, the front panel toggle switch must first be flipped to the down position, then the above sequence can be repeated with the indicator on, but the ON/clr button must be pressed to reset the zero position reading to .0.

Note: DO NOT Change resolutions and inch/millimeter functions on the digital indicator. They are not active, but doing so can result in requiring the indicator to be reset. The digital indicator will turn off if not used for 10 minutes when battery powered. A power adapter is available to provide continuous power if preferred




H-1240DA Parts
(Parts which do not appear on the H-1240 Drawing on the following page.)

Penetration Needles and Cones

renetration Needles and Cones							
Application	Specification	Model	Description				
		H-1280	Standard hardened stainless steel needle, 40-45mm exposed needle length. Wt. 2.5g.				
Dita con in a con one attention of	ASTM D5; AASH- TO T49, IP49; ASA Std., A37.1; Fed Spec. SS.R. 406C, Meth. 214.01	H-1300	Same as H-1280. Certified to ASTM accuracy by independent laboratory. Wt. 2.5g.				
Bituminous materials		H-1302	Same as H-1290. Certified to ASTM accuracy by independent laboratory. Wt. 2.5g.				
		H-1290	Long hardened stainless steel needle, 50-55mm exposed needle length. Wt. 2.5g.				
Waxes with 250 or less penetration	ASTM D1321	H-1310	Hardened stainless steel wax penetration needles with tapered point, blunt tip of truncated cone. Ferrule is approx. 3.2mm dia. Wt. 2.5g.				
		H-1317	Same as H-1310. Certified to ASTM accuracy by independent laboratory.				
Battery paste	N/A	H-1255	Hardened stainless steel tip with special plunger. Total wt. 60g ± .050g.				
Joint sealant for asphalt & concrete pavements	ASTM D5329	H-1320	Resilience ball penetration tool. Total wt. 27.5g.				
Grease-testing	ASTM D217,	H-2520	Hollow 90° brass cone, highly polished stainless steel tip. Removable nut and stem. Wt. 102.5g.				
penetrometers	D937 ASA Std. Z11.3	H-2522	Same as H-2520, completely made of stainless steel. Wt. 102.5g.				
For all 90° cones		H-2525	Stainless steel replacement tip, nut and stem.				
Grease-testing penetrometers	ASTM D2884	H-2524	Magnesium cone and plunger, same dim. as H-2520. Total wt. cone and plunger 30.0g.				
Food products and paste paints	U.S. Dept. of Agriculture	H-2529	Aluminum. Same dimension as H-2520. Total wt. 35g.				
(shortenings, margarine, butter, etc.) Measuring firmness of solid and plasticized fats	AOCS Cc 16-60	H-1270	10° aluminum cone, 3.2mm ferrule, 0.8mm stainless steel blunt tip. Overall length 106mm. Wt. 45g.				
(i.e., recovery of used grease) Small obtainable samples	ASTM D1403	H-2519	1/4 scale. (Not considered a substitute for full-size cone specified in ASTM D217.) Wt. 9.38g.				





H-1240 Parts Drawing (Refer to previous page for items that differ from the H-1240 Drawing above.)

Warranty

Humboldt Mfg. Co. warrants its products to be free from defects in material or workmanship. The exclusive remedy for this warranty is Humboldt Mfg. Co., factory replacement of any part or parts of such product, for the warranty of this product please refer to Humboldt Mfg. Co. catalog on Terms and Conditions of Sale. The purchaser is responsible for the transportation charges. Humboldt Mfg. Co. shall not be responsible under this warranty if the goods have been improperly maintained, installed, operated or the goods have been altered or modified so as to adversely affect the operation, use performance or durability or so as to change their intended use. The Humboldt Mfg. Co. liability under the warranty contained in this clause is limited to the repair or replacement of defective goods and making good, defective workmanship.

Humboldt Mfg. Co.

875 Tollgate Road Elgin, Illinois 60123 U.S.A. U.S.A. Toll Free: 1.800.544.7220

Voice: 1.708.456.6300

Fax: 1.708.456.0137 Email: hmc@humboldtmfg.com

Testing Equipment for



Construction Materials

HUMBOLDT

www.humboldtmfg.com